



SIGNAL CONDITIONER

LVDT/RVDT

MACRO EAZY-CAL™ LVC-4000

Overview

The EAZY-CAL™ LVC-4000 is a standalone signal conditioner, supporting a wide range of AC LVDTs, RVDTs, and VR half-bridges, while providing several choices of voltage, current, and digital RS-485 outputs. Push-button calibration offers intuitive operation as compared to signal conditioners with span and offset trim pots. Fault conditions, such as a wire break on LVDT/RVDT connections, are indicated by blinking LEDs, fault condition error output, and Error Flag Open Collector signal (see manual for details). The LVC-4000 operates from a 9-30V DC power supply and is housed in a polyamide DIN rail-mounted enclosure. Calibration instructions, terminal functions, LVDT connection diagram and DIP switch functions are printed on the side panels for convenience.

Synchronization to other signal conditioners is accomplished by a daisy chain connection to a synchronization bus. One unit will assume the Master function based on DIP switch priority setting. If a fault should occur, the next highest priority unit will take over as Master.

With the use of the RS-485 port, a host computer is able to retrieve measurement data, receive operational status, perform remote calibration, and perform hot swap reconfiguration.

Features

- ◆ Push-button or RS-485 command auto-calibration
- ◆ Analog voltage or current loop output
- ◆ Digital RS-485 interface
- ◆ Supports standard AC LVDTs, RVDTs, and VR half-bridge sensors
- ◆ Master/slave excitation synchronization
- ◆ DIN-rail mountable
- ◆ Color-coded terminal blocks

User Selectable Features

- ◆ 0-5V DC, 0-10V DC, 0.5-4.5V DC, $\pm 5V$ DC, $\pm 10V$ DC or 4-20 mA output
- ◆ 1.5V_{rms} or 3.0V_{rms} sensor excitation
- ◆ 2.5, 5, 7.5, or 10 kHz excitation frequency

Environmental Data

Operating Temperature	-20 to 75°C (0 to 165°F)
Temperature Sensitivity	<0.02% of FSO/°C (<0.01% of FSO/°F)
EMC Compliance	Emissions: EN55011:2007 Immunity: EN61000-4-2:2009 EN61000-4-4:2004 EN61000-4-6:2009 EN61000-4-3:2010+A2:2010



Electrical Data

Power Input	9-30V DC (90 mA max. @ 24V DC)	Output Non-Linearity	≤±0.1% full scale output
Sensor Excitation	3.0V _{rms} (1.5V _{rms} selectable)	Output Voltage Ripple	1 mV _{rms} max. (2.5 kHz excitation, no filter) 2 mV _{rms} max. (10 kHz excitation, no filter)
Sensor Excitation Frequency	2.5 kHz, 5 kHz, 7.5 kHz, or 10 kHz	Output Current Ripple	10 μA_{rms} max. (2.5 kHz excitation, no filter) 20 μA_{rms} max. (10 kHz excitation, no filter)
Input Sensitivity Range	55 mV _{rms} to 5.5 V _{rms} full scale input produces full scale DC output	Frequency Response (-3dB)	500 Hz max.
Full Scale Outputs	0-5V DC, 0-10V DC, 0.5-4.5V DC, $\pm 5V$ DC, $\pm 10V$ DC or 4-20 mA output		


SIGNAL CONDITIONER

EAZY-CAL™ LVC-4000

Connection Diagrams



Dimensions



NORTH AMERICA

Measurement Specialties, Inc.,
a TE Connectivity Company
Phone +1-800-522-6752
Email: customercare.pens@te.com

TE.com/sensorsolutions

Measurement Specialties, Inc., a TE Connectivity company.

Accustar, American Sensor Technologies, AST, ATEXIS, DEUTSCH, IdentiCal, TruBlue, KPSI, Krystal Bond, Microfused, UltraStable, Measurement Specialties, MEAS, Schaevitz, TE, and the TE connectivity (logo) are trademarks of the TE Connectivity Ltd. family of companies. Other logos, product and company names mentioned herein may be trademarks of their respective owners.

The information given herein, including drawings, illustrations and schematics which are intended for illustration purposes only, is believed to be reliable. However, TE Connectivity makes no warranties as to its accuracy or completeness and disclaims any liability in connection with its use. TE Connectivity's obligations shall only be as set forth in TE Connectivity's Standard Terms and Conditions of Sale for this product and in no case will TE Connectivity be liable for any incidental, indirect or consequential damages arising out of the sale, resale, use or misuse of the product. Users of TE Connectivity products should make their own evaluation to determine the suitability of each such product for the specific application.