1000 Watt Medical



Features

- 5 x 9.51 x 1.61 inches
- Approval to EN60601 Edition 3.1
- Dual Fusing
- Current Sharing Option
- Peak Power Capability
- Standard IEC60601-1-2: 2014 (4th Edition)
- 5 Vdc Stand by
- In built 12 V fan output
- Power Good / Power Fail Signal
- Suitable for BF application
- Lesser than 1U high
- Having high voltage output range up to 58VDC
- N+1 redundant power supply
- · Single wire current sharing
- Built in OR-ing diode / FET (- R suffix)

	Electrical Specifications	S
Input Voltage	85-264 VAC/120-390 VDC, Universal	
Input Frequency	47-63 Hz	
Input Current	120 VAC: 11 A max.	240 VAC: 5.5 A max.
Input Protection	F16A/250 V in Live & Neutral both	
No Load Power	Typ 3W over entire input range with main output kept off using Remote ON/OFF	
Inrush Current	240 VAC: 25 A max.	
Leakage Current	400 μA @ 240 VAC / 50 Hz	Touch Current : < 100 μA
Efficiency	120 VAC: 88% Typical 240 VAC: 93%	
Hold-up Time	120 VAC: 8 ms	240 VAC: 8 ms
Power Factor	120 VAC: 0.98	240 VAC: 0.95
Output Power	1000W Fan Cooled, Peak 1200W for 1m	nS
Line Regulation	+/-0.5%	
Load Regulation	+/-1%	
Transient Response	< 10%, 50% to 100% load change, 50 Hz, 50% duty cycle, 0.1 A/μs, recovery time < 5 ms	
Rise Time	<100 ms	
Set Point Tolerance	+/-1%	
Output Adjustability	+/-3%	
Over Current Protection	110% Typ, HiccUp Type, Autorecovery	
Over Voltage Protection	114%, Latch Type, AC Power to be recy	cled for recovery
Short Circuit Protection	Latch Type, AC Power to be recycled for	recovery
Over Temperature Protection	130-140°C primary heat sink, autorecov	rery
Current Share	Upto 3 Supplies can be connected in pa	rallel (optional)
Switching Frequency	PFC converter:Variable, 85 kHz typical	
	Resonant converter: Variable, 100 kHz ty	/pical
Operating Temperature	-40 to +70°C, refer derating curve	
Storage Temperature	−40 to +85°C	
Relative Humidity	95% Rh, noncondensing	
Altitude	Operating: 16,000 ft.; Nonoperating: 40,000 ft.	
MTBF	3.37m Hours, Telcordia -SR332-issue 3	
Isolation Voltage	Input to Output 4245 VAC, Input to Earth 1625 VAC, Output to Earth 1500 VAC	
Cooling	Fan Cooled : 1000W	

Model Number	Туре	Voltage	Max. Load	Min. Load	Ripple ¹
MVPS1000-1012	Fan Cooled	12V	41.67 A	0.0 A	2%
MVPS1000-1015	Fan Cooled	15 V	41.67 A	0.0 A	2%
MVPS1000-1024	Fan Cooled	24 V	41.67 A	0.0 A	2%
MVPS1000-1030	Fan Cooled	30 V	33.33 A	0.0 A	2%
MVPS1000-1048	Fan Cooled	48 V	20.83 A	0.0 A	2%
MVPS1000-1058	Fan Cooled	58 V	17.24 A	0.0 A	2%

Pin Connections				
J1	1	AC LINE		
	2	NEUTRAL		
	3	EARTH		
J2	J2-A	+VE		
	J2-B	-VE		
J3	Pin 1	GND		
	Pin 2	5V AUX		
	Pin 3	PGPF		
	Pin 4	VS -		
	Pin 5	VS+		
	Pin 6	GND		
	Pin 7	RMT		
	Pin 8	CL2		
	Pin 9	CL1		
	Pin 10	LS		
J10, J11 (Fan Output)	Pin 1	+VE		
	Pin 2	-VE		

Notes

- 1. For Ripple measurement minimum output power requirement is 25 W.


 Ripple is peak to peak with 20 MHz bandwidth and 10 μF (Electrolytic capacitor) in parallel with a 0.1 μF capacitor at rated line voltage and load ranges.
- 2. Combined output power of main output, fan supply and standby supply shall not exceed max. power rating.
- 3. Standby output voltage 5 V/1.5 A with tolerance including set point accuracy, line and load regulation is +/-10%. Ripple and noise is less than 5%.
- 4. Specifications are for nominal input voltage, 25°C unless otherwise stated.
- 5. PSU is supplied with J3, pin-6 and pin-7 shorted to enable main output without remote on/off feature.
- 6. Fan supply output voltage is 12V/500mA with regulation band+/-30 % and Ripple is less than 10%. To get 12V Fan supply output voltage, minimum 10 % load on Main output voltage is required.

	Mechanical Specifications		
AC Input Connector (J1)	TE Connectivity: NC6-P107-03		
DC Output Connector (J2)	6-32 inches Screw Pan HD		
	Mating: Designed to accept Ring Tongue Terminal AMP: 8-31886-1,		
	wherein one 16 AWG(max) wire can be crimp	ed.	
	Note: One Ring Tongue Terminal with 16 AWG is	s recommended for current upto 11A only.	
	Use multiple tongue terminals with wire for more current.		
Signal Connector (J3)	Molex: 22-23-2101		
	Mating: 22-01-2107; Pins: 08-50-0113		
J10, J11 (Fan Output)	Make: TE Connectivity AMP Connectors		
	Description: CONN HEADER VERT 2POS 2.54N	1M	
	MPN : 640456-2		
	Mating: 3-641535-2 / TE Connectivity AMP Connectors OR		
	0022013027 / MOLEX with crimping 08-50-01	14 / MOLEX	
Dimensions	5.0 x 9.51 x 1.61 inches		
W. L.	(127 x 241.5 x 41 mm)		
Weight	1.3 kg		
	EMC		
Parameter	Conditions/Description	Criteria	
Conducted Emissions	EN 55011-B,CISPR22-B, FCC PART15-B	Class B	
Radiated Emissions	EN 55011	Class A (Class B with External king c	
		K5B RC 25x12x15-M or equivalent)	
Input Current Harmonics	EN 61000-3-2	Class A	
Voltage Fluctuation and Flicker	EN 61000-3-3	Complies	
ESD Immunity	EN 61000-4-2	A	
Radiated Field Immunity	EN 61000-4-3	A	
Electrical Fast Transient Immunity	EN 61000-4-4	A	
Surge Immunity	EN 61000-4-5	A	
Conducted Immunity	EN 61000-4-6	A	
Magnetic Field Immunity	EN 61000-4-8	A	
Voltage dips, interruptions	EN 61000-4-11	A & B	
	Safety		
CE Mark	Complies with LVD Directive		
Approval Agency	Nemko, UL, C-UL		
Safety Standard(s)	EN60601-1, IEC 60601-1 (ed.3),ANSI/AAMI ES 60601-1,		
	CSA C22.2 No. 60601-1		
Safety File Number(s)	UL Certificate No : 2019-02-21-E173812		
	CB Test Certificate No : NO105338		
	Nemko Certificate No : P19223365		


	Signal(s)	
Power Good / Power Fail Signal	Power Good: Is a TTL signal which goes high after main output reaches 90% of its set value.	
The delay is 0.1 s to 0.5 s		
	Power Fail: The same signal goes low at least 1ms before main output falls to 90% of set	
value at AC Power off		
	Signal 90%Vout T 100-500mS > 1mS Power Good Power Fail	
Remote Sense	Compensates for 200 mV drop	
Remote on/off	Pin 6 & Pin 7 of J3 can be used for Remote on/off.	
	Shorting Pin 6 to Pin 7 enables main output while keeping the pins open disables main output	
	Note: - Provision of Inhibit Remote ON/OFF is available. +5V at Pin 7 will switch off the main output	
OCP limit set	Pin 8 & Pin 9 of J3 must be shorted	



Derating Curve w.r.t Input




Derating Curve for 12 V




De-rate between 30-50 °C @ 0.833% per °C De-rate above 50 °C @ 2.5% per °C

5







Installtion instruction for current sharing:

During the installation and setup of parallel supplies in a system it is important that a single remote sense point be used for all the supplies. The remote sense voltage between the supplies must be adjusted to within 1% to ensure the supplies are inside the 1% capture window. If the supplies are not initially adjusted inside the capture window the supplies will not current share satisfactorily.

Set-Up Procedures:

- 1. Connect load cables to the outputs of each supply.
- 2. Connect the remote sense lines to the load in twisted style. (A common remote sense point must be used for all the supplies in parallel).
- 3. Connect all the "LS" signal(Pin 10) on the J3 connector between the supplies.
- 4. Adjust remote sense voltage of each supply to within 1% of rated output voltage or readjust to required set point. (Adjustment to be done with all other parallel supplies off).
- 5. Current sharing between the supplies can be verified by monitoring the output current of each supply with a hall effect DC current probe. The supplies should share to within 10% of the total load current.
 - The maximum recommended power output for three units in parallel would be 2700W.
- 6. The current share circuit has a capture window voltage of +/- 1% of the rated output voltage. If the output remote sense voltage of one of the supplies is adjusted outside the 1% window the supplies will not current share satisfactorily.

CURRENT SHARING BLOCK DIAGRAM

